

**Test Report**

No.T32120310515SN

Date: Oct 25, 2021

Page 1 of 2

SAVEWO LIMITED  
1/F., 266-270 TEXACO ROAD, TSUEN WAN, N.T., HONG KONG

The following samples were submitted and identified by/on behalf of the client as:

3DMEOW

Case No. : CA321203122489  
 Lot No. / Batch Code : NOT PROVIDED  
 Sample Description : WHITE MASK  
 Style/ Item No. : 3DMEOW  
 Supplier : SAVEWO LIMITED  
 Manufacturer : SAVEWO LIMITED  
 Country of Origin : HONG KONG  
 Sample Receiving Date : OCT 20, 2021  
 Testing Period : OCT 20, 2021 – OCT 22, 2021

Test Requested	Conclusion
Resistance to penetration by synthetic blood (ASTM F1862/F1862M-17)	PASS

\*\*\*\*\* FOR FURTHER DETAILS, PLEASE REFER TO THE FOLLOWING PAGE(S) \*\*\*\*\*

Signed for and on behalf of  
SGS Hong Kong Ltd.



Au Kam Chi, Gigi  
Technical Manager

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.



# Test Report

No.T32120310515SN

Date: Oct 25, 2021

Page 2 of 2

## Test Results:

### Resistance to penetration by synthetic blood ASTM F1862/F1862M-17


Test Side : Outside  
 Pre-Conditioning : Minimum of 4 hours at 21±5°C and 85±5% R.H.  
 Test Condition : 21±5°C and 85±10% R.H.  
 Test Pressure : 160 mmHg  
 No of Test Specimen Tested : 32  
 No of Test Specimen Passed : 30

Test Specimen #	Synthetic Blood Penetration
1-13, 15-21, 23-32	None Seen
14, 22	Yes

#### Note:

An acceptable quality limit of 4,0 % is met for a single sampling plan when 29 or more of the 32 tested specimens show "pass" results.

#### Sample Photo:



SGS authenticate the photo on original report only

\*\*\* End of Report \*\*\*

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.